Certificate No.：L0372-200117

財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

This is to certify that

TUNG PEI INDUSTRIAL CO., LTD.

Length Calibration Laboratory

No.7, Songjiang N. Rd., Zhongli Dist., Taoyuan City 320, Taiwan

is accredited in respect of laboratory

Accreditation Criteria	:	ISO/IEC 17025:2017;CNS 17025:2018
Accreditation Number	:	0372
Originally Accredited	:	January 01, 1999
Effective Period	:	January 20, 2020 to January 19, 2023
Accredited Scope	:	Calibration Field, see described in the Appendix

Chung-Lin Wang

Chung-Lin Wang
President, Taiwan Accreditation Foundation
Date: January 17, 2020

The Appendix forms an integral part of this Certificate, which shall be invalid when use without the Appendix

Certificate No.：L0372-200117

財團法人全國認證基金會
Taiwan Accreditation Foundation

Accreditation Number ：0372

Laboratory Head ：LIN, Ho-Nein

Length

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA1001 Gauge Block (CNS 8092 /112 pcs) (Steel)	P.T.W B-1 Grade 00	Gauge Block Calibration Procedure (Document No.:Q10IT901A)	0.5	mm	100	mm	(20±1.5) ℃ /L:mm, length of gauge block	$[(65)^2+(1.7\ L)^2]^{1/2}$	nm

Note: Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.

Approval Signatory

Approval Signatory	Scope
LIN, Ho-Nein	KA1001
CHEN, Kuan-Min	KA1001

(Null Below)

The Appendix forms an integral part of this Certificate, which shall be invalid when use without the Appendix